

The Reuse of Single-Use Devices

**FDA Proposed Strategy:
Concept in Development**

Beginning of Practice

- Reuse of Reusable Devices Started in 1960s
- Advent of Single Use Only Devices in Early 1980s
- Economics is Driving Force for Reprocessing
- Hospital v. Third-Party Reprocessing
- Most Frequently Reprocessed Single-Use Devices

FDA's Position Historically

- Reprocessing in Hospitals/Clinics (Compliance Policy Guide 300.500)
- Any Person Engaged in Single Use Device Reprocessing is a “Manufacturer”
- Premarket Submissions Have Not Been Requested

FDA's Position Historically

(continued)

- Requirements of 3rd Party Reprocessing Firms:
 - Device Registration and Listing, 21 CFR, Part 807
 - Good Manufacturing Practice (GMP) Inspection, 21 CFR, Part 820
 - Medical Device Reporting, 21 CFR, Part 803
 - General Labeling Requirements, 21 CFR, Part 801
- Reuse Policy Documents & Correspondence on FDA Web Page (www.fda.gov/reuse)

Simple Solutions?

- One voice in the debate suggests calling for identical regulatory controls for reprocessing as for OEMs - call for 510(k)s and PMAs
- An opposing voice suggests we leave General Controls in place as sufficient: Registration and Listing, GMP (Quality System Requirements), Labeling, and Medical Device Reporting
- Neither approach is satisfactory

Problems to Solve

- Lack of evidence of public health problems does not mean that the current practice is safe and effective
- This system inside hospitals and in third parties has grown over time with FDA tacit acceptance
- Reuse is basically a problem of economics and ethics: both are outside of FDA mandate!

Some Guiding Principles

- Capitalize on what we do best: understanding of regulatory control and devices
- Our constraints suggest the importance of partnering/outside leveraging: show leadership but do not solve all by ourselves
- Do not let the perfect serve as the enemy of the good

Regulatory Strategy by Risk

Product Risk Category	Regulatory Requirements	Enforcement Date
“High-Risk” Products	R & L; Premarket submissions or Cease reprocessing	Immediate enforcement action within 6 months
“Moderate-Risk” Products	R & L; Collect postmarket data on S & E; Declare conformance or file 510 (k)	R & L – 6 mos Submissions to FDA within 2 years
“Low-Risk” Products	R & L	Within 1 year

How do we get there?

- Four committees working rapidly:
 - **Steering (Kessler)**
 - **Policy (Ng)**
 - **Categorization (Zimmerman)**
 - **Research (Merritt)**
- Milestones
 - **Strategy paper: October**
 - **Teleconference: Nov. 10**
 - **Public Meeting: Dec 14**
 - **Product specific Guidance : February, 2000**
 - **Enforcement: Spring, 2000**

Possible Roles for Standards

- Standardized methods for cleaning, disinfection, sterilization
- Verification of sterility
- Measurement of endotoxins
- Verification of device performance after reprocessing

Critical Premarket Issues

- How to establish device specifications to ensure device is (as) safe and effective
- How to detect changes to devices by OEM and the need for revalidation
- Ability to perform thorough process definition and validation studies given facility and sterilizer limitations

Vision for the Future

Current Reality

- Widespread practice with little data on safety or effectiveness
- Single use labels not clearly meaningful
- Single use labels don't identify vulnerabilities
- Patients are not informed - experimentation?

Future Vision

- FDA regulatory approach will be RISK and SCIENCE based
- Single use labels will have clinical relevance
- Single use labels will identify vulnerabilities
- Horizontal and vertical standards critical
- Leverage outside parties